

In Reference to Case No(s):

4201/1572J

Contract Laboratory Program  
REGIONAL/LABORATORY COMMUNICATION SYSTEM

Telephone Record Log

Date of Call: 25 July 85

Laboratory Name: California Analytical

Lab Contact: Karin Yee / Charles Soderqvist

Region: 10

Regional Contact: Andrew Hafferty

#3  
Gary Beltz

Call Initiated By: Laboratory ☐ Region ☒

In reference to data for the following sample number(s):

1291  
all SAS samples 1572J (2,4-D and 2,4,5-TP analyses)  
240 + 245 T LLDs - why a change

Summary of Questions/Issues Discussed:

1. Sample data for 1291 reported 1290 on run data - dates & times etc are OK - appears to be mislabeled
2. Was SAS analysis done with Method Esters
3. Why are some LLDs elevated

Summary of Resolution:

1. Will send corrected sheets
2. Yes - Method Esters were made a GC/MS
3. LLDs elevated due to A. dilution effect  
B. noisy blank  
C. matrix interference  
Data is acceptable - OK

Signature

25 July 85  
Date

Distribution: (1)Lab Copy, (2)Region Copy, (3)SMO Copy

USEPA SF



1624815